

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
 Data File : VV023615.D
 Acq On : 18 Nov 2021 19:40
 Operator : SY/MD
 Sample : M4694-09DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H4672DL

Quant Time: Nov 19 04:09:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 03:51:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	110436	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	106855	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	49074	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	33966	4.910	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	98.200%	
7) Chloroethane-d5	1.568	69	28720	5.093	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	101.800%	
11) 1,1-Dichloroethene-d2	2.108	63	48610	3.753	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	75.000%	
20) 2-Butanone-d5	3.895	46	48928	41.050	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	82.100%	
24) Chloroform-d	4.349	84	68832	4.668	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.400%	
26) 1,2-Dichloroethane-d4	5.037	65	32893	4.961	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.200%	
32) Benzene-d6	5.050	84	127623	4.655	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.000%	
36) 1,2-Dichloropropane-d6	6.072	67	38304	4.746	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	95.000%	
41) Toluene-d8	7.317	98	108762	4.233	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.600%	
43) trans-1,3-Dichloroprop...	7.625	79	14120	4.614	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.200%	
46) 2-Hexanone-d5	8.091	63	49997	44.404	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.800%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26999	4.652	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.000%	
66) 1,2-Dichlorobenzene-d4	11.625	152	42419	5.191	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.800%	
Target Compounds						
					Qvalue	
5) Vinyl chloride	1.310	62	8069	0.882	ug/L	89
17) Methyl tert-butyl Ether	2.770	73	8146	0.562	ug/L	98
18) trans-1,2-Dichloroethene	2.764	96	5826	0.720	ug/L	96
22) cis-1,2-Dichloroethene	3.912	96	59950	7.695	ug/L #	90
34) Trichloroethene	5.918	95	19981	2.516	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
Data File : VV023615.D
Acq On : 18 Nov 2021 19:40
Operator : SY/MD
Sample : M4694-09DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H4672DL

Quant Time: Nov 19 04:09:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 03:51:44 2021
Response via : Initial Calibration

