

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111821\
 Data File : VW023620.D
 Acq On : 18 Nov 2021 21:39
 Operator : SY/MD
 Sample : M4706-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AA4

Quant Time: Nov 19 04:10:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 03:51:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	111794	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	111534	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	49400	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	33357	4.763	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.200%
7) Chloroethane-d5	1.568	69	27529	4.823	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.400%
11) 1,1-Dichloroethene-d2	2.108	63	47026	3.587	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.800%
20) 2-Butanone-d5	3.896	46	57581	47.723	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.440%
24) Chloroform-d	4.352	84	54292	3.638	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.800%
26) 1,2-Dichloroethane-d4	5.034	65	31415	4.681	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.050	84	122649	4.286	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.800%
36) 1,2-Dichloropropane-d6	6.069	67	37964	4.507	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.200%
41) Toluene-d8	7.317	98	108055	4.029	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.600%
43) trans-1,3-Dichloroprop...	7.625	79	13181	4.126	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.600%
46) 2-Hexanone-d5	8.092	63	47639	40.534	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.060%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	25057	4.136	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	40389	4.910	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	1992	0.215	ug/L	99
13) Acetone	2.185	43	3631	4.925	ug/L	97
25) Chloroform	4.371	83	12943	0.878	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
Data File : VV023620.D
Acq On : 18 Nov 2021 21:39
Operator : SY/MD
Sample : M4706-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AA4

Quant Time: Nov 19 04:10:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 03:51:44 2021
Response via : Initial Calibration

