

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111821\
 Data File : VW023624.D
 Acq On : 18 Nov 2021 23:15
 Operator : SY/MD
 Sample : M4706-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AB3

Quant Time: Nov 19 04:10:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 03:51:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	106953	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	107402	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	46937	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	30126	4.496	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.000%
7) Chloroethane-d5	1.568	69	26084	4.777	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.600%
11) 1,1-Dichloroethene-d2	2.108	63	44160	3.521	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.400%
20) 2-Butanone-d5	3.905	46	37998	32.918	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	65.840%
24) Chloroform-d	4.349	84	56118	3.930	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.600%
26) 1,2-Dichloroethane-d4	5.034	65	30227	4.707	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%
32) Benzene-d6	5.050	84	115988	4.209	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.072	67	35200	4.339	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.317	98	97575	3.779	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.600%
43) trans-1,3-Dichloroprop...	7.625	79	12696	4.127	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.600%
46) 2-Hexanone-d5	8.091	63	45311	40.037	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.080%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	24463	4.193	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	38826	4.968	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	5323	0.600	ug/L	99
13) Acetone	2.185	43	1888	2.677	ug/L	60
25) Chloroform	4.378	83	8269	0.586	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
Data File : VV023624.D
Acq On : 18 Nov 2021 23:15
Operator : SY/MD
Sample : M4706-07
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AB3

Quant Time: Nov 19 04:10:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 03:51:44 2021
Response via : Initial Calibration

