

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111821\
 Data File : VW023628.D
 Acq On : 19 Nov 2021 00:50
 Operator : SY/MD
 Sample : M4706-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AB9

Quant Time: Nov 19 04:11:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 03:51:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	109664	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	108832	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	49212	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	32569	4.741	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.800%
7) Chloroethane-d5	1.568	69	26247	4.688	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.800%
11) 1,1-Dichloroethene-d2	2.108	63	45367	3.527	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.600%
20) 2-Butanone-d5	3.899	46	51030	43.115	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.220%
24) Chloroform-d	4.349	84	61004	4.167	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
26) 1,2-Dichloroethane-d4	5.034	65	32340	4.912	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.050	84	120799	4.326	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.600%
36) 1,2-Dichloropropane-d6	6.072	67	36351	4.422	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.400%
41) Toluene-d8	7.317	98	102448	3.915	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.400%
43) trans-1,3-Dichloroprop...	7.625	79	13207	4.237	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.092	63	47335	41.276	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.560%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	25528	4.318	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	46758	5.706	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.200%
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	2804	0.308	ug/L	85
13) Acetone	2.188	43	2582	3.570	ug/L	76
25) Chloroform	4.375	83	7452	0.515	ug/L	95
42) Toluene	7.397	91	5135	0.158	ug/L	86
47) Tetrachloroethene	7.979	164	1001	0.143	ug/L #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
Data File : VV023628.D
Acq On : 19 Nov 2021 00:50
Operator : SY/MD
Sample : M4706-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AB9

Quant Time: Nov 19 04:11:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 03:51:44 2021
Response via : Initial Calibration

