

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW111821\
 Data File : VW023629.D
 Acq On : 19 Nov 2021 01:14
 Operator : SY/MD
 Sample : M4706-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 B0AA2

Quant Time: Nov 19 04:11:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Nov 19 03:51:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	110202	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	110486	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	49016	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	31666	4.587	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	91.800%	
7) Chloroethane-d5	1.568	69	26190	4.655	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.000%	
11) 1,1-Dichloroethene-d2	2.108	63	43442	3.361	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	67.200%	
20) 2-Butanone-d5	3.889	46	59040	49.639	ug/L	-0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	99.280%	
24) Chloroform-d	4.349	84	62721	4.263	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.200%	
26) 1,2-Dichloroethane-d4	5.034	65	30839	4.661	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
32) Benzene-d6	5.050	84	116336	4.104	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.000%	
36) 1,2-Dichloropropane-d6	6.069	67	36215	4.340	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	86.800%	
41) Toluene-d8	7.317	98	100340	3.777	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	75.600%	
43) trans-1,3-Dichloroprop...	7.625	79	12280	3.881	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	77.600%	
46) 2-Hexanone-d5	8.088	63	44052	37.838	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	75.680%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	24760	4.126	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	82.600%	
66) 1,2-Dichlorobenzene-d4	11.625	152	38663	4.737	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.800%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	3491	0.382	ug/L	98
13) Acetone	2.182	43	4481	6.166	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111821\
Data File : VV023629.D
Acq On : 19 Nov 2021 01:14
Operator : SY/MD
Sample : M4706-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
B0AA2

Quant Time: Nov 19 04:11:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR110421WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Nov 19 03:51:44 2021
Response via : Initial Calibration

