

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111823\
 Data File : VV033003.D
 Acq On : 18 Nov 2023 13:45
 Operator : SY/MD
 Sample : VV1118WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 20 01:19:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 20 01:19:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	372295	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	375367	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	189121	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	116701	37.063	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	74.120%	
7) Chloroethane-d5	1.529	69	89669	34.371	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	68.740%#	
11) 1,1-Dichloroethene-d2	2.056	65	46341	33.779	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	67.560%	
21) 2-Butanone-d5	3.780	46	136471	101.113	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	101.110%	
24) Chloroform-d	4.246	84	245613	45.152	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.300%	
26) 1,2-Dichloroethane-d4	4.941	65	152395	46.042	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	92.080%	
32) Benzene-d6	4.957	84	483568	47.234	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.460%	
36) 1,2-Dichloropropane-d6	5.985	67	140819	47.321	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	94.640%	
41) Toluene-d8	7.243	98	433941	45.974	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.940%	
43) trans-1,3-Dichloroprop...	7.551	79	72161	45.677	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	91.360%	
47) 2-Hexanone-d5	8.024	63	106968	98.118	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	98.120%	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	187091	42.918	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	85.840%	
66) 1,2-Dichlorobenzene-d4	11.561	152	182852	49.758	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	99.520%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111823\
Data File : VV033003.D
Acq On : 18 Nov 2023 13:45
Operator : SY/MD
Sample : VV1118WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 20 01:19:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 20 01:19:10 2023
Response via : Initial Calibration

