

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111823\
 Data File : VV033009.D
 Acq On : 18 Nov 2023 17:32
 Operator : SY/MD
 Sample : VIBLK248
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK248

Quant Time: Nov 20 01:20:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Nov 20 01:19:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	292855	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	297761	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	149169	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	81723	32.994	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.980%
7) Chloroethane-d5	1.529	69	67437	32.861	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	65.720%#
11) 1,1-Dichloroethene-d2	2.059	65	32320	29.949	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.900%#
21) 2-Butanone-d5	3.780	46	128126	120.680	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	120.680%
24) Chloroform-d	4.249	84	211186	49.354	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.700%
26) 1,2-Dichloroethane-d4	4.940	65	143361	55.062	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.120%
32) Benzene-d6	4.960	84	399577	49.202	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.400%
36) 1,2-Dichloropropane-d6	5.989	67	117546	49.795	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.600%
41) Toluene-d8	7.243	98	359454	48.008	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.020%
43) trans-1,3-Dichloroprop...	7.554	79	63510	50.679	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.360%
47) 2-Hexanone-d5	8.024	63	105477	121.966	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	121.970%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	174296	50.404	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	158879	54.814	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV111823\
Data File : VV033009.D
Acq On : 18 Nov 2023 17:32
Operator : SY/MD
Sample : VIBLK248
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK248

Quant Time: Nov 20 01:20:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM111323WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Nov 20 01:19:10 2023
Response via : Initial Calibration

