

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111824\
 Data File : VV038162.D
 Acq On : 18 Nov 2024 09:59
 Operator : SY/MD
 Sample : VV1118WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Nov 19 00:10:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	9545	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	9421	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4591	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.281	65	5003	0.506	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	102.000%
4) 1,2-Dichloroethane-d4	4.970	65	3148	0.526	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	5.996	67	3727	0.488	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	98.000%
8) Toluene-d8	7.254	98	9449	0.484	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3014	0.494	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	98.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111824\
Data File : VV038162.D
Acq On : 18 Nov 2024 09:59
Operator : SY/MD
Sample : VV1118WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Nov 19 00:10:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

