

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111824\
 Data File : VV038163.D
 Acq On : 18 Nov 2024 10:42
 Operator : SY/MD
 Sample : P4847-09
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK002

Quant Time: Nov 19 00:10:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Nov 15 00:15:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	8951	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	8972	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	4453	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	5445	0.587	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 118.000%	
4) 1,2-Dichloroethane-d4	4.969	65	3264	0.581	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 116.000%	
7) 1,2-Dichloropropane-d6	5.995	67	3934	0.541	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 108.000%	
8) Toluene-d8	7.253	98	9926	0.533	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3116	0.536	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 108.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111824\
Data File : VV038163.D
Acq On : 18 Nov 2024 10:42
Operator : SY/MD
Sample : P4847-09
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Quant Time: Nov 19 00:10:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM111424.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Nov 15 00:15:53 2024
Response via : Initial Calibration

