

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111924\
 Data File : VV038177.D
 Acq On : 19 Nov 2024 14:53
 Operator : SY/MD
 Sample : VV1119WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK256

Quant Time: Nov 19 22:47:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 19 22:47:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	331718	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	340701	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	152175	5.000	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	118977	4.679	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.529	69	100063	4.865	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.400%
11) 1,1-Dichloroethene-d2	2.053	65	53105	4.854	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.000%
20) 2-Butanone-d5	3.796	46	223082	43.859	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.720%
24) Chloroform-d	4.240	84	227808	4.617	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
26) 1,2-Dichloroethane-d4	4.934	65	107068	4.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
32) Benzene-d6	4.950	84	434568	4.935	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.800%
36) 1,2-Dichloropropane-d6	5.979	67	132399	4.886	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.236	98	364587	4.475	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	7.555	79	45783	4.545	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.021	63	181629	46.382	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.760%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	104146	4.528	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	11.554	152	140642	5.424	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV111924\
Data File : VV038177.D
Acq On : 19 Nov 2024 14:53
Operator : SY/MD
Sample : VV1119WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK256

Quant Time: Nov 19 22:47:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 19 22:47:22 2024
Response via : Initial Calibration

