

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112022\
 Data File : VV029239.D
 Acq On : 20 Nov 2022 21:43
 Operator : SY/MD
 Sample : VV1120WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK273

Quant Time: Nov 21 03:37:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Nov 21 03:37:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	333676	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	304961	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	139578	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	105280	5.175	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 103.600%		
7) Chloroethane-d5	1.568	69	84547	4.728	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 94.600%		
11) 1,1-Dichloroethene-d2	2.105	63	129730	3.499	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 70.000%		
20) 2-Butanone-d5	3.909	46	109662	32.204	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 64.400%		
24) Chloroform-d	4.343	84	178489	4.274	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 85.400%		
26) 1,2-Dichloroethane-d4	5.027	65	82274	4.378	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 87.600%		
32) Benzene-d6	5.047	84	381581	4.192	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 83.800%		
36) 1,2-Dichloropropane-d6	6.066	67	111057	4.369	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 87.400%		
41) Toluene-d8	7.310	98	336328	4.105	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 82.000%		
43) trans-1,3-Dichloroprop...	7.622	79	30561	3.686	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 73.800%		
46) 2-Hexanone-d5	8.088	63	141572	40.008	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 80.020%		
56) 1,1,2,2-Tetrachloroeth...	10.211	84	76125	4.320	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 86.400%		
66) 1,2-Dichlorobenzene-d4	11.616	152	118546	4.747	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 95.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112022\
Data File : VV029239.D
Acq On : 20 Nov 2022 21:43
Operator : SY/MD
Sample : VV1120WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK273

Quant Time: Nov 21 03:37:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111122WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Nov 21 03:37:08 2022
Response via : Initial Calibration

