

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112024\
 Data File : VV038208.D
 Acq On : 20 Nov 2024 12:20
 Operator : SY/MD
 Sample : P4882-12DL 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3H9DL

Quant Time: Nov 21 01:47:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 21 01:41:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	214436	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	224591	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	96692	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	74673	4.543	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.800%
7) Chloroethane-d5	1.526	69	63699	4.791	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.800%
11) 1,1-Dichloroethene-d2	2.044	65	33962	4.802	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.000%
20) 2-Butanone-d5	3.805	46	200417	60.954	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.900%
24) Chloroform-d	4.240	84	170323	5.340	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.800%
26) 1,2-Dichloroethane-d4	4.937	65	84499	5.904	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.000%
32) Benzene-d6	4.950	84	316181	5.447	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.000%
36) 1,2-Dichloropropane-d6	5.982	67	104036	5.825	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	116.400%
41) Toluene-d8	7.239	98	254905	4.746	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
43) trans-1,3-Dichloroprop...	7.558	79	35399	5.331	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	106.600%
46) 2-Hexanone-d5	8.034	63	174909	67.757	ug/L	0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	135.520%#
56) 1,1,2,2-Tetrachloroeth...	10.149	84	89100	5.876	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.600%
66) 1,2-Dichlorobenzene-d4	11.558	152	101585	6.166	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	123.400%#
Target Compounds						
					Qvalue	
6) Bromomethane	1.484	94	1183	0.153	ug/L	88
9) Trichlorofluoromethane	1.696	101	376878	13.378	ug/L	100
10) 1,1,2-Trichloro-1,2,2-...	2.050	101	6166	0.384	ug/L	93
12) 1,1-Dichloroethene	2.056	96	3884	0.254	ug/L #	1
16) Methylene chloride	2.439	84	18428	1.055	ug/L	98
19) 1,1-Dichloroethane	3.101	63	30412	1.038	ug/L	97
22) cis-1,2-Dichloroethene	3.818	96	3455	0.207	ug/L	96
29) 1,1,1-Trichloroethane	4.500	97	16274	0.604	ug/L	97
34) Trichloroethene	5.834	95	21806	1.268	ug/L	94
47) Tetrachloroethene	7.902	164	12789	0.925	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112024\
Data File : VV038208.D
Acq On : 20 Nov 2024 12:20
Operator : SY/MD
Sample : P4882-12DL 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3H9DL

Quant Time: Nov 21 01:47:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 21 01:41:56 2024
Response via : Initial Calibration

