

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW112222\
 Data File : VW029292.D
 Acq On : 23 Nov 2022 03:29
 Operator : SY/MD
 Sample : N5652-14DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH1M3DL

Quant Time: Nov 23 04:49:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112222WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 23 01:46:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	276931	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	246841	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	115915	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	82327	3.117	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	62.400%	
7) Chloroethane-d5	1.564	69	76047	3.678	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	73.600%	
11) 1,1-Dichloroethene-d2	2.104	63	105939	2.354	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	47.000%#	
20) 2-Butanone-d5	3.902	46	221045	56.688	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	113.380%	
24) Chloroform-d	4.339	84	150041	4.005	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	80.200%	
26) 1,2-Dichloroethane-d4	5.027	65	77504	4.374	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.400%	
32) Benzene-d6	5.043	84	299523	3.722	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	74.400%	
36) 1,2-Dichloropropane-d6	6.063	67	107192	4.216	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	84.400%	
41) Toluene-d8	7.310	98	244760	3.375	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	67.600%#	
43) trans-1,3-Dichloroprop...	7.616	79	31346	3.702	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.000%	
46) 2-Hexanone-d5	8.088	63	95115	62.668	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	125.340%	
56) 1,1,2,2-Tetrachloroeth...	10.207	84	63679	5.037	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	100.800%	
66) 1,2-Dichlorobenzene-d4	11.615	152	88394	3.953	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	79.000%#	
Target Compounds						
					Qvalue	
16) Methylene chloride	2.503	84	17380	0.540	ug/L	96
18) trans-1,2-Dichloroethene	2.757	96	8007	0.386	ug/L	100
22) cis-1,2-Dichloroethene	3.902	96	20037	0.884	ug/L	97
34) Trichloroethene	5.905	95	168622	7.129	ug/L	98
47) Tetrachloroethene	7.969	164	6799	0.438	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV112222\
Data File : VV029292.D
Acq On : 23 Nov 2022 03:29
Operator : SY/MD
Sample : N5652-14DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH1M3DL

Quant Time: Nov 23 04:49:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR112222WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 23 01:46:18 2022
Response via : Initial Calibration

