

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112224\
 Data File : VV038284.D
 Acq On : 22 Nov 2024 21:55
 Operator : SY/MD
 Sample : P4883-20
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 23 02:05:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 23 02:01:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	200119	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	220449	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	87018	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	70199	4.577	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.600%
7) Chloroethane-d5	1.529	69	68559	5.526	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.600%
11) 1,1-Dichloroethene-d2	2.053	65	32642	4.945	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	99.000%
20) 2-Butanone-d5	3.793	46	191914	62.544	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.080%
24) Chloroform-d	4.243	84	165720	5.567	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.400%
26) 1,2-Dichloroethane-d4	4.937	65	80538	6.030	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	120.600%
32) Benzene-d6	4.953	84	284874	5.000	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%
36) 1,2-Dichloropropane-d6	5.982	67	97315	5.551	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.000%
41) Toluene-d8	7.239	98	210740	3.997	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.000%
43) trans-1,3-Dichloroprop...	7.558	79	29809	4.573	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.021	63	128911	50.877	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.760%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	84639	5.687	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	85237	5.749	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.000%
Target Compounds					Qvalue	
47) Tetrachloroethene	7.902	164	4139	0.305	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112224\
Data File : VV038284.D
Acq On : 22 Nov 2024 21:55
Operator : SY/MD
Sample : P4883-20
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 23 02:05:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 23 02:01:29 2024
Response via : Initial Calibration

