

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112324\
 Data File : VV038307.D
 Acq On : 23 Nov 2024 18:49
 Operator : SY/MD
 Sample : P4884-19DL 100X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3P4DL

Quant Time: Nov 23 23:02:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Nov 23 02:01:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	206567	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	221294	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	90491	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	57068	3.604	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.000%
7) Chloroethane-d5	1.529	69	60928	4.757	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.053	65	26234	3.850	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.000%
20) 2-Butanone-d5	3.793	46	183101	57.809	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.620%
24) Chloroform-d	4.240	84	153693	5.002	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%
26) 1,2-Dichloroethane-d4	4.937	65	80000	5.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.000%
32) Benzene-d6	4.954	84	261550	4.573	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	5.982	67	91919	5.223	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.400%
41) Toluene-d8	7.240	98	197720	3.736	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.800%
43) trans-1,3-Dichloroprop...	7.558	79	28771	4.397	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.000%
46) 2-Hexanone-d5	8.021	63	155821	61.262	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.520%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	81860	5.479	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.600%
66) 1,2-Dichlorobenzene-d4	11.555	152	84580	5.486	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.800%
Target Compounds					Qvalue	
22) cis-1,2-Dichloroethene	3.815	96	2672	0.166	ug/L	88
34) Trichloroethene	5.831	95	40743	2.404	ug/L	95
47) Tetrachloroethene	7.896	164	276696	20.321	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112324\
Data File : VV038307.D
Acq On : 23 Nov 2024 18:49
Operator : SY/MD
Sample : P4884-19DL 100X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3P4DL

Quant Time: Nov 23 23:02:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Nov 23 02:01:29 2024
Response via : Initial Calibration

