

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112524\
 Data File : VV038354.D
 Acq On : 25 Nov 2024 19:42
 Operator : SY/MD
 Sample : P4884-21
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 26 00:48:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Nov 26 00:39:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	208946	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	232664	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	91071	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	60179	3.758	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.200%
7) Chloroethane-d5	1.529	69	57626	4.448	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.053	65	27997	4.062	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.200%
20) 2-Butanone-d5	3.789	46	152527	47.608	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.220%
24) Chloroform-d	4.240	84	144571	4.651	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	4.937	65	64790	4.646	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.000%
32) Benzene-d6	4.953	84	235057	3.909	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.200%
36) 1,2-Dichloropropane-d6	5.982	67	80386	4.344	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.239	98	176633	3.174	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.400%#
43) trans-1,3-Dichloroprop...	7.558	79	22857	3.323	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.400%
46) 2-Hexanone-d5	8.021	63	101227	37.853	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.700%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	71145	4.529	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	11.554	152	71368	4.599	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112524\
Data File : VV038354.D
Acq On : 25 Nov 2024 19:42
Operator : SY/MD
Sample : P4884-21
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 26 00:48:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Nov 26 00:39:03 2024
Response via : Initial Calibration

