

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112624\
 Data File : VV038375.D
 Acq On : 26 Nov 2024 19:31
 Operator : SY/MD
 Sample : P4930-10
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 27 00:26:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 27 00:19:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	196094	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	221284	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	86725	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	54384	3.618	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.400%
7) Chloroethane-d5	1.526	69	54893	4.515	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.200%
11) 1,1-Dichloroethene-d2	2.053	65	26573	4.108	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.200%
20) 2-Butanone-d5	3.789	46	148987	49.551	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.100%
24) Chloroform-d	4.240	84	139167	4.771	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.400%
26) 1,2-Dichloroethane-d4	4.937	65	67399	5.150	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	4.953	84	225922	3.950	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.000%
36) 1,2-Dichloropropane-d6	5.982	67	80759	4.589	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.800%
41) Toluene-d8	7.239	98	171165	3.234	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.600%#
43) trans-1,3-Dichloroprop...	7.561	79	22032	3.367	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.400%
46) 2-Hexanone-d5	8.021	63	92180	36.243	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.480%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	71064	4.757	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.200%
66) 1,2-Dichlorobenzene-d4	11.554	152	68505	4.636	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112624\
Data File : VV038375.D
Acq On : 26 Nov 2024 19:31
Operator : SY/MD
Sample : P4930-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 27 00:26:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 27 00:19:46 2024
Response via : Initial Calibration

