

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112724\
 Data File : VV038381.D
 Acq On : 27 Nov 2024 17:46
 Operator : SY/MD
 Sample : P4884-21
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Nov 27 18:05:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Nov 27 00:19:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	205106	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	228337	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	91990	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	81248	5.168	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.400%
7) Chloroethane-d5	1.529	69	74351	5.847	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.000%
11) 1,1-Dichloroethene-d2	2.053	65	36093	5.335	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	106.800%
20) 2-Butanone-d5	3.802	46	176851	56.234	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.460%
24) Chloroform-d	4.236	84	170579	5.591	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.800%
26) 1,2-Dichloroethane-d4	4.934	65	81489	5.952	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.000%
32) Benzene-d6	4.947	84	292461	4.956	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	5.979	67	99868	5.499	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.000%
41) Toluene-d8	7.236	98	221442	4.055	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
43) trans-1,3-Dichloroprop...	7.555	79	29435	4.360	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.200%
46) 2-Hexanone-d5	8.021	63	113707	43.326	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.660%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	82690	5.364	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.200%
66) 1,2-Dichlorobenzene-d4	11.555	152	87952	5.612	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112724\
Data File : VV038381.D
Acq On : 27 Nov 2024 17:46
Operator : SY/MD
Sample : P4884-21
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Nov 27 18:05:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Nov 27 00:19:46 2024
Response via : Initial Calibration

