

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112724\
 Data File : VV038379.D
 Acq On : 27 Nov 2024 11:18
 Operator : SY/MD
 Sample : VV1127WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK266

Quant Time: Nov 28 00:55:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Nov 28 00:29:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	212725	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	236572	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	96570	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	61147	3.750	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.000%
7) Chloroethane-d5	1.529	69	62247	4.720	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.400%
11) 1,1-Dichloroethene-d2	2.053	65	28731	4.095	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.800%
20) 2-Butanone-d5	3.802	46	159492	48.898	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.800%
24) Chloroform-d	4.240	84	161460	5.102	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.000%
26) 1,2-Dichloroethane-d4	4.934	65	73310	5.163	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.200%
32) Benzene-d6	4.950	84	261511	4.277	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	5.979	67	90918	4.832	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.236	98	203422	3.596	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.000%
43) trans-1,3-Dichloroprop...	7.561	79	27533	3.936	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.800%
46) 2-Hexanone-d5	8.024	63	100869	37.096	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	74.200%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	73787	4.620	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.400%
66) 1,2-Dichlorobenzene-d4	11.554	152	81774	4.970	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV112724\
Data File : VV038379.D
Acq On : 27 Nov 2024 11:18
Operator : SY/MD
Sample : VV1127WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK266

Quant Time: Nov 28 00:55:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Nov 28 00:29:54 2024
Response via : Initial Calibration

