

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120424\
 Data File : VV038408.D
 Acq On : 04 Dec 2024 11:27
 Operator : SY/MD
 Sample : VV1204WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK215

Quant Time: Dec 05 12:47:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Dec 03 21:43:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7428	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	7461	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3615	0.500	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	3867	0.564	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	112.000%
4) 1,2-Dichloroethane-d4	4.969	65	3087	0.586	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	118.000%
7) 1,2-Dichloropropane-d6	5.995	67	3757	0.537	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.000%
8) Toluene-d8	7.253	98	7981	0.487	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2997	0.552	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	110.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120424\
Data File : VV038408.D
Acq On : 04 Dec 2024 11:27
Operator : SY/MD
Sample : VV1204WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK215

Quant Time: Dec 05 12:47:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM120324.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Dec 03 21:43:44 2024
Response via : Initial Calibration

