

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
 Data File : VV038420.D
 Acq On : 05 Dec 2024 16:26
 Operator : SY/MD
 Sample : VV1206WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK267

Quant Time: Dec 06 05:40:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 06 05:39:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	224366	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	249155	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	101734	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	103634	6.026	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.600%
7) Chloroethane-d5	1.529	69	86119	6.191	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	123.800%
11) 1,1-Dichloroethene-d2	2.050	65	45791	6.187	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	123.800%
20) 2-Butanone-d5	3.815	46	173017	50.292	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.580%
24) Chloroform-d	4.240	84	188361	5.644	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.800%
26) 1,2-Dichloroethane-d4	4.937	65	88235	5.892	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.800%
32) Benzene-d6	4.950	84	344012	5.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.800%
36) 1,2-Dichloropropane-d6	5.982	67	109486	5.525	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.600%
41) Toluene-d8	7.236	98	269838	4.529	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.600%
43) trans-1,3-Dichloroprop...	7.558	79	35640	4.838	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.800%
46) 2-Hexanone-d5	8.024	63	123612	43.165	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.320%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	84320	5.013	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.200%
66) 1,2-Dichlorobenzene-d4	11.554	152	100316	5.787	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
Data File : VV038420.D
Acq On : 05 Dec 2024 16:26
Operator : SY/MD
Sample : VV1206WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK267

Quant Time: Dec 06 05:40:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 06 05:39:27 2024
Response via : Initial Calibration

