

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
 Data File : VV038426.D
 Acq On : 05 Dec 2024 19:17
 Operator : SY/MD
 Sample : P5116-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E1PJ8

Quant Time: Dec 06 05:41:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 06 05:39:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	208106	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	235905	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	99222	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	98413	6.170	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	123.400%
7) Chloroethane-d5	1.529	69	85541	6.630	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	132.600%#
11) 1,1-Dichloroethene-d2	2.053	65	42200	6.148	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	123.000%
20) 2-Butanone-d5	3.793	46	171586	53.773	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.540%
24) Chloroform-d	4.243	84	180030	5.816	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.400%
26) 1,2-Dichloroethane-d4	4.937	65	87156	6.275	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	125.400%
32) Benzene-d6	4.953	84	330168	5.415	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.400%
36) 1,2-Dichloropropane-d6	5.982	67	103772	5.531	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.600%
41) Toluene-d8	7.239	98	255955	4.537	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
43) trans-1,3-Dichloroprop...	7.558	79	32124	4.605	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.200%
46) 2-Hexanone-d5	8.021	63	130782	48.233	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.460%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	84792	5.324	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.400%
66) 1,2-Dichlorobenzene-d4	11.554	152	96343	5.699	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
Data File : VV038426.D
Acq On : 05 Dec 2024 19:17
Operator : SY/MD
Sample : P5116-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E1PJ8

Quant Time: Dec 06 05:41:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 06 05:39:27 2024
Response via : Initial Calibration

