

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
 Data File : VV038428.D
 Acq On : 05 Dec 2024 20:06
 Operator : SY/MD
 Sample : P5111-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG451

Quant Time: Dec 06 05:42:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 06 05:39:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	214744	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	241853	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	93087	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	104755	6.364	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	127.200%
7) Chloroethane-d5	1.529	69	85485	6.421	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	128.400%
11) 1,1-Dichloroethene-d2	2.053	65	44010	6.213	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	124.200%
20) 2-Butanone-d5	3.796	46	176619	53.639	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.280%
24) Chloroform-d	4.240	84	180896	5.663	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.200%
26) 1,2-Dichloroethane-d4	4.937	65	84510	5.896	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.000%
32) Benzene-d6	4.953	84	332573	5.320	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.400%
36) 1,2-Dichloropropane-d6	5.982	67	105850	5.503	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.000%
41) Toluene-d8	7.239	98	253172	4.377	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
43) trans-1,3-Dichloroprop...	7.558	79	27921	3.904	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.000%
46) 2-Hexanone-d5	8.021	63	119999	43.168	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.340%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	84751	5.190	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	92792	5.851	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
Data File : VV038428.D
Acq On : 05 Dec 2024 20:06
Operator : SY/MD
Sample : P5111-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG451

Quant Time: Dec 06 05:42:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 06 05:39:27 2024
Response via : Initial Calibration

