

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
 Data File : VV038429.D
 Acq On : 05 Dec 2024 20:30
 Operator : SY/MD
 Sample : P5111-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG452

Quant Time: Dec 06 05:42:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 06 05:39:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	200262	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.780	117	228879	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	88860	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	99659	6.493	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 129.800%	
7) Chloroethane-d5	1.529	69	85910	6.919	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 138.400%#	
11) 1,1-Dichloroethene-d2	2.053	65	41516	6.285	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 125.800%#	
20) 2-Butanone-d5	3.799	46	183858	59.876	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	= 119.760%	
24) Chloroform-d	4.240	84	184273	6.186	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 123.800%	
26) 1,2-Dichloroethane-d4	4.937	65	88585	6.627	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 132.600%#	
32) Benzene-d6	4.953	84	336100	5.682	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 113.600%	
36) 1,2-Dichloropropane-d6	5.982	67	108769	5.975	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 119.600%	
41) Toluene-d8	7.236	98	255682	4.671	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 93.400%	
43) trans-1,3-Dichloroprop...	7.558	79	31766	4.694	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	= 93.800%	
46) 2-Hexanone-d5	8.021	63	125213	47.597	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 95.200%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	86301	5.585	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 111.600%	
66) 1,2-Dichlorobenzene-d4	11.554	152	95611	6.315	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 126.400%#	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120624\
Data File : VV038429.D
Acq On : 05 Dec 2024 20:30
Operator : SY/MD
Sample : P5111-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG452

Quant Time: Dec 06 05:42:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 06 05:39:27 2024
Response via : Initial Calibration

