

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120724\
 Data File : VV038438.D
 Acq On : 06 Dec 2024 14:53
 Operator : SY/MD
 Sample : P5111-08
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG457

Quant Time: Dec 06 21:36:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 06 21:32:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	226432	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.779	117	247642	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	100019	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	95431	5.499	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.000%
7) Chloroethane-d5	1.529	69	82628	5.886	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.800%
11) 1,1-Dichloroethene-d2	2.053	65	42344	5.670	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	113.400%
20) 2-Butanone-d5	3.799	46	182663	52.611	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.220%
24) Chloroform-d	4.239	84	181423	5.386	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.800%
26) 1,2-Dichloroethane-d4	4.937	65	86120	5.698	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	114.000%
32) Benzene-d6	4.950	84	327417	5.115	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
36) 1,2-Dichloropropane-d6	5.982	67	104549	5.308	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.200%
41) Toluene-d8	7.236	98	259511	4.382	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
43) trans-1,3-Dichloroprop...	7.558	79	32786	4.478	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.600%
46) 2-Hexanone-d5	8.021	63	127555	44.814	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.620%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	85623	5.121	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.400%
66) 1,2-Dichlorobenzene-d4	11.554	152	97821	5.740	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV120724\
Data File : VV038438.D
Acq On : 06 Dec 2024 14:53
Operator : SY/MD
Sample : P5111-08
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG457

Quant Time: Dec 06 21:36:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR111924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 06 21:32:44 2024
Response via : Initial Calibration

