

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121920\
 Data File : VV019762.D
 Acq On : 20 Dec 2020 02:24
 Operator : SY/MD
 Sample : L5131-13
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Dec 21 07:12:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 21 06:23:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	702806	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	662100	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	295713	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	230112	45.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.32%
7) Chloroethane-d5	1.57	69	189848	47.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.52%
11) 1,1-Dichloroethene-d2	2.11	63	339374	36.28	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.56%
21) 2-Butanone-d5	3.90	46	326113	119.05	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.05%
24) Chloroform-d	4.35	84	426791	46.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.98%
26) 1,2-Dichloroethane-d4	5.04	65	295050	49.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.42%
32) Benzene-d6	5.05	84	861904	48.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.62%
36) 1,2-Dichloropropane-d6	6.07	67	277650	50.19	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.38%
41) Toluene-d8	7.32	98	752935	47.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.72%
43) trans-1,3-Dichloropropene-	7.62	79	122977	39.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.18%
47) 2-Hexanone-d5	8.09	63	197074	90.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.78%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	333421	44.56	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.12%
64) 1,2-Dichlorobenzene-d4	11.63	152	271313	49.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.16%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121920\
Data File : VV019762.D
Acq On : 20 Dec 2020 02:24
Operator : SY/MD
Sample : L5131-13
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Dec 21 07:12:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 21 06:23:42 2020
Response via : Initial Calibration

