

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV121920\
 Data File : VV019745.D
 Acq On : 19 Dec 2020 18:13
 Operator : SY/MD
 Sample : L5101-07
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Dec 21 06:09:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Dec 21 03:36:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	703101	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	656887	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.25	152	299477	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	240784	47.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.52%
7) Chloroethane-d5	1.57	69	196181	49.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.68%
11) 1,1-Dichloroethene-d2	2.11	63	352058	37.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.24%
21) 2-Butanone-d5	3.89	46	355859	129.86	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	129.86%
24) Chloroform-d	4.35	84	440157	48.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.88%
26) 1,2-Dichloroethane-d4	5.04	65	306726	51.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.30%
32) Benzene-d6	5.05	84	891904	50.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.82%
36) 1,2-Dichloropropane-d6	6.07	67	286834	52.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.52%
41) Toluene-d8	7.32	98	784694	49.75	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.50%
43) trans-1,3-Dichloropropene-	7.62	79	132682	42.50	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.00%
47) 2-Hexanone-d5	8.09	63	219277	101.80	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.80%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	352452	47.48	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.96%
64) 1,2-Dichlorobenzene-d4	11.63	152	285557	51.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.02%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV121920\
Data File : VV019745.D
Acq On : 19 Dec 2020 18:13
Operator : SY/MD
Sample : L5101-07
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Dec 21 06:09:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Dec 21 03:36:59 2020
Response via : Initial Calibration

