

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV122120\
 Data File : VV019773.D
 Acq On : 21 Dec 2020 13:07
 Operator : SY/MD
 Sample : L5098-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Dec 22 00:47:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 22 00:24:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.62	114	703672	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.86	117	657182	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.26	152	300979	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	201602	39.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.90%
7) Chloroethane-d5	1.57	69	171308	43.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.10%
11) 1,1-Dichloroethene-d2	2.11	63	308664	32.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.92%
21) 2-Butanone-d5	3.89	46	340683	124.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	124.22%
24) Chloroform-d	4.35	84	410058	45.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.18%
26) 1,2-Dichloroethane-d4	5.04	65	286460	48.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.40%
32) Benzene-d6	5.05	84	812332	46.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.70%
36) 1,2-Dichloropropane-d6	6.07	67	263801	48.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.10%
41) Toluene-d8	7.32	98	710099	45.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	7.62	79	125711	40.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.50%
47) 2-Hexanone-d5	8.09	63	215594	100.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.05%
57) 1,1,2,2-Tetrachloroethane-	10.22	84	335529	45.18	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.36%
64) 1,2-Dichlorobenzene-d4	11.63	152	261270	46.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.88%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV122120\
Data File : VV019773.D
Acq On : 21 Dec 2020 13:07
Operator : SY/MD
Sample : L5098-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Dec 22 00:47:55 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM121020WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 22 00:24:34 2020
Response via : Initial Calibration

