

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW010319\
 Data File : VW008020.D
 Acq On : 03 Jan 2019 14:56
 Operator : SY/AP
 Sample : K1017-01
 Misc : 4.94G/5ML/MSVOA W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-1

Quant Time: Jan 04 07:15:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 03 05:03:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	64545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	105739	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	78861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	24872	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	38813	54.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.74%	
35) Dibromofluoromethane	7.88	113	34025	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
50) Toluene-d8	10.33	98	123545	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
62) 4-Bromofluorobenzene	12.62	95	33381	34.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.62%	
Target Compounds						
					Qvalue	
16) Acetone	4.12	43	7821	70.772	ug/l	99
17) Carbon Disulfide	4.37	76	2022	1.021	ug/l	97
25) 2-Butanone	7.18	43	1710	13.133	ug/l #	86
52) Toluene	10.39	92	3202	1.763	ug/l	96
95) Naphthalene	15.38	128	79185	87.829	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010319\
Data File : VW008020.D
Acq On : 03 Jan 2019 14:56
Operator : SY/AP
Sample : K1017-01
Misc : 4.94G/5ML/MSVOA W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-1

Quant Time: Jan 04 07:15:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 03 05:03:55 2019
Response via : Initial Calibration

