

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW010319\
 Data File : VW008023.D
 Acq On : 03 Jan 2019 16:17
 Operator : SY/AP
 Sample : K1017-13
 Misc : 5.27G/5ML/MSVOA W/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TP-4

Manual Integrations
 APPROVED

MMDadoda
 1/4/2019 12:02:17 PM

Quant Time: Jan 04 07:23:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 03 05:03:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	64206	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	108608	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	84469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	31836	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	39068	55.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.02%	
35) Dibromofluoromethane	7.88	113	34455	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
50) Toluene-d8	10.32	98	127237	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	12.62	95	38901	39.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.98%	
Target Compounds						
16) Acetone	4.12	43	6240	56.764	ug/l	97
20) Methylene Chloride	4.92	84	2606	1.541	ug/l #	84
25) 2-Butanone	7.19	43	1154m	8.910	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010319\
Data File : VW008023.D
Acq On : 03 Jan 2019 16:17
Operator : SY/AP
Sample : K1017-13
Misc : 5.27G/5ML/MSVOA W/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TP-4

Manual Integrations
APPROVED

MMDadoda
1/4/2019 12:02:17 PM

Quant Time: Jan 04 07:23:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 03 05:03:55 2019
Response via : Initial Calibration

