

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW010320\
 Data File : VW014513.D
 Acq On : 03 Jan 2020 21:42
 Operator : SY/VA
 Sample : K6482-07
 Misc : 4.85G/5ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 MC-123119-4-5-GRAB-B4

Quant Time: Jan 06 01:24:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 03 16:07:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	725758	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	1149143	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	908029	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	327618	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	364254	57.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.80%	
35) Dibromofluoromethane	7.88	113	374575	57.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.06%	
50) Toluene-d8	10.32	98	1421896	51.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.76%	
62) 4-Bromofluorobenzene	12.62	95	394145	42.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.80%	
Target Compounds						
8) Diethyl Ether	3.67	74	4262	1.160	ug/l	79
16) Acetone	4.12	43	15049	8.399	ug/l	96
20) Methylene Chloride	4.92	84	93234	9.176	ug/l	94
68) m/p-Xylenes	11.83	106	18443	1.437	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010320\
Data File : VW014513.D
Acq On : 03 Jan 2020 21:42
Operator : SY/VA
Sample : K6482-07
Misc : 4.85G/5ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
MC-123119-4-5-GRAB-B4

Quant Time: Jan 06 01:24:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 03 16:07:20 2020
Response via : Initial Calibration

