

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW010419\
 Data File : VW008055.D
 Acq On : 04 Jan 2019 18:35
 Operator : SY/AP
 Sample : K1044-01
 Misc : 5.03G/5ML/MSVOA W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CPSB-08(10-15)RE

Quant Time: Jan 05 03:16:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 03 05:03:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	72270	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.85	114	121910	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	90072	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	36582	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	42672	53.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.76%	
35) Dibromofluoromethane	7.89	113	37396	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
50) Toluene-d8	10.32	98	139954	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
62) 4-Bromofluorobenzene	12.62	95	44159	39.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.88%	
Target Compounds						
					Qvalue	
16) Acetone	4.15	43	5805	46.914	ug/l	93
17) Carbon Disulfide	4.38	76	5055	2.281	ug/l #	84
20) Methylene Chloride	4.91	84	2778	1.327	ug/l	98
25) 2-Butanone	7.20	43	1663	11.407	ug/l #	81
39) Methylcyclohexane	9.34	83	2841	1.895	ug/l #	90
85) sec-Butylbenzene	13.39	105	13069	4.337	ug/l #	55
95) Naphthalene	15.38	128	33983	25.627	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010419\
Data File : VW008055.D
Acq On : 04 Jan 2019 18:35
Operator : SY/AP
Sample : K1044-01
Misc : 5.03G/5ML/MSVOA W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CPSB-08(10-15)RE

Quant Time: Jan 05 03:16:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 03 05:03:55 2019
Response via : Initial Calibration

