


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010419\
Data File : VW008038.D
Acq On : 04 Jan 2019 09:49
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
Title : SW846 8260
Last Update : Thu Jan 03 05:03:55 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	11424	PASS
75	95	30	60	53.6	32344	PASS
95	95	100	100	100.0	60389	PASS
96	95	5	9	6.7	4045	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	73.6	44442	PASS
175	174	5	9	6.1	2695	PASS
176	174	95	101	100.4	44640	PASS
177	176	5	9	6.4	2838	PASS