

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010622\
 Data File : VW021760.D
 Acq On : 06 Jan 2022 16:20
 Operator : SY/VA
 Sample : VIBLK577
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK577

Quant Time: Jan 07 00:22:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Jan 04 00:22:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	140979	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	142173	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	79528	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	47754	17.180	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.720%
7) Chloroethane-d5	2.885	69	37966	20.279	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.120%
11) 1,1-Dichloroethene-d2	4.013	63	56855	13.977	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.920%
21) 2-Butanone-d5	7.086	46	16410	40.352	ug/L	0.01
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.700%
24) Chloroform-d	7.653	84	91636	20.808	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.240%
26) 1,2-Dichloroethane-d4	8.311	65	49236	21.173	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	84.680%
32) Benzene-d6	8.275	84	189832	22.487	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	89.960%
36) 1,2-Dichloropropane-d6	9.275	67	50453	22.031	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.120%
41) Toluene-d8	10.323	98	184328	21.915	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.640%
43) trans-1,3-Dichloroprop...	10.579	79	21059	18.898	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.600%
47) 2-Hexanone-d5	10.921	63	14942	43.022	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.040%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	40977	21.704	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.800%
66) 1,2-Dichlorobenzene-d4	13.853	152	69673	22.910	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW010622\
Data File : VW021760.D
Acq On : 06 Jan 2022 16:20
Operator : SY/VA
Sample : VIBLK577
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK577

Quant Time: Jan 07 00:22:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM122821SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Jan 04 00:22:49 2022
Response via : Initial Calibration

