

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW010719\
 Data File : VW008063.D
 Acq On : 07 Jan 2019 12:15
 Operator : SY/AP
 Sample : VW0107SBL01
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0107SBL01

Quant Time: Jan 08 03:54:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 03 05:03:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	70841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	115216	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	97536	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.57	152	42885	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	40106	51.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.38%	
35) Dibromofluoromethane	7.88	113	35693	49.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.28%	
50) Toluene-d8	10.32	98	135021	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
62) 4-Bromofluorobenzene	12.62	95	48687	46.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.18%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010719\
Data File : VW008063.D
Acq On : 07 Jan 2019 12:15
Operator : SY/AP
Sample : VW0107SBL01
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0107SBL01

Quant Time: Jan 08 03:54:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
Quant Title : SW846 8260
QLast Update : Thu Jan 03 05:03:55 2019
Response via : Initial Calibration

