


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010719\
Data File : VW008061.D
Acq On : 07 Jan 2019 09:43
Operator : SY/AP
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010219S.M
Title : SW846 8260
Last Update : Thu Jan 03 05:03:55 2019



AutoFind: Scans 1757, 1758, 1759; Background Corrected with Scan 1749

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.4	12196	PASS
75	95	30	60	51.6	34237	PASS
95	95	100	100	100.0	66341	PASS
96	95	5	9	6.7	4461	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	74.9	49688	PASS
175	174	5	9	6.3	3108	PASS
176	174	95	101	97.0	48210	PASS
177	176	5	9	6.6	3196	PASS