

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW010820\
 Data File : VW014555.D
 Acq On : 08 Jan 2020 12:42
 Operator : SY/VA
 Sample : K6476-04
 Misc : 7.17G/5ML/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WP022SB454_191227-24-26

Quant Time: Jan 08 13:25:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 03 16:07:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	677038	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	1044337	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	827235	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	333525	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	275868	46.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.20%	
35) Dibromofluoromethane	7.88	113	295819	49.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.12%	
50) Toluene-d8	10.32	98	1193263	47.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.82%	
62) 4-Bromofluorobenzene	12.62	95	348973	41.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.62%	
Target Compounds						
16) Acetone	4.12	43	29501	17.650	ug/l	97
17) Carbon Disulfide	4.37	76	23739	1.169	ug/l	99
51) 4-Methyl-2-Pentanone	10.20	43	6374	1.480	ug/l	95
58) 2-Chloroethyl Vinyl ether	9.92	63	5310	2.666	ug/l	95
65) Chlorobenzene	11.66	112	20586	1.215	ug/l #	86
85) sec-Butylbenzene	13.38	105	49450	1.986	ug/l #	55
89) n-Butylbenzene	13.82	91	57356	2.721	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW010820\
Data File : VW014555.D
Acq On : 08 Jan 2020 12:42
Operator : SY/VA
Sample : K6476-04
Misc : 7.17G/5ML/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WP022SB454_191227-24-26

Quant Time: Jan 08 13:25:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 03 16:07:20 2020
Response via : Initial Calibration

