

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW010820\
 Data File : VW014558.D
 Acq On : 08 Jan 2020 14:01
 Operator : SY/VA
 Sample : K6476-05
 Misc : 10.00G/5ML/MSVOA_W/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WP022SB455_191227-12-14

Manual Integrations
 APPROVED

apatel
 1/10/2020 2:41:29 PM

Quant Time: Jan 08 14:32:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 03 16:07:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	665549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	1051063	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	839883	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	327747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	266576	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
35) Dibromofluoromethane	7.88	113	288512	48.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.06%	
50) Toluene-d8	10.32	98	1159529	46.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.50%	
62) 4-Bromofluorobenzene	12.62	95	341462	40.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.32%	
Target Compounds						
16) Acetone	4.12	43	19941	12.136	ug/l	90
20) Methylene Chloride	4.92	84	41804	1.948	ug/l	95
65) Chlorobenzene	11.66	112	19292	1.121	ug/l	99
87) 1,3-Dichlorobenzene	13.50	146	13402	1.234	ug/l	94
88) 1,4-Dichlorobenzene	13.58	146	14229m	1.327	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW010820\
Data File : VW014558.D
Acq On : 08 Jan 2020 14:01
Operator : SY/VA
Sample : K6476-05
Misc : 10.00G/5ML/MSVOA_W/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WP022SB455_191227-12-14

Manual Integrations
APPROVED

apatel
1/10/2020 2:41:29 PM

Quant Time: Jan 08 14:32:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 03 16:07:20 2020
Response via : Initial Calibration

