

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW010919\
 Data File : VW008132.D
 Acq On : 10 Jan 2019 05:52
 Operator : SY/AP
 Sample : K1087-06RE
 Misc : 5.86G/5ML/MSVOA W/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 MH-447-7.0-8.0RE

Quant Time: Jan 10 08:21:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 04:40:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	57672	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	107934	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	96186	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	42432	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	38577	59.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.04%	
35) Dibromofluoromethane	7.88	113	34432	54.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.40%	
50) Toluene-d8	10.32	98	140239	53.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.18%	
62) 4-Bromofluorobenzene	12.62	95	44534	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
Target Compounds						
16) Acetone	4.14	43	3853	13.022	ug/l	99
20) Methylene Chloride	4.92	84	6768	6.414	ug/l #	93
52) Toluene	10.39	92	22584	11.727	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010919\
Data File : VW008132.D
Acq On : 10 Jan 2019 05:52
Operator : SY/AP
Sample : K1087-06RE
Misc : 5.86G/5ML/MSVOA W/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
MH-447-7.0-8.0RE

Quant Time: Jan 10 08:21:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 04:40:48 2019
Response via : Initial Calibration

