

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_W\DATA\VW010820\
 Data File : VW014563.D
 Acq On : 08 Jan 2020 16:49
 Operator : SY/VA
 Sample : L1056-04
 Misc : 6.82G/5ML/MSVOA W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Jan 10 03:25:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 03 16:07:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	568168	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	877392	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	697031	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	279283	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	250742	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
35) Dibromofluoromethane	7.88	113	236017	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
50) Toluene-d8	10.32	98	874180	41.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.54%	
62) 4-Bromofluorobenzene	12.62	95	272599	38.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.80%	
Target Compounds						
4) Vinyl Chloride	2.36	62	2307143	399.473	ug/l	98
12) 1,1-Dichloroethene	4.04	96	9863	1.744	ug/l	92
16) Acetone	4.12	43	59932	42.727	ug/l	97
18) Methyl Acetate	4.67	43	69221	20.433	ug/l	99
21) trans-1,2-Dichloroethene	5.42	96	12218	2.035	ug/l	90
24) 1,1-Dichloroethane	6.21	63	43140	4.229	ug/l	97
27) cis-1,2-Dichloroethene	7.16	96	3509455	556.313	ug/l	93
42) 1,2-Dichloroethane	8.39	62	11245	1.811	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_W\DATA\VW010820\
Data File : VW014563.D
Acq On : 08 Jan 2020 16:49
Operator : SY/VA
Sample : L1056-04
Misc : 6.82G/5ML/MSVOA W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Jan 10 03:25:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Quant Title : SW846 8260
Last Update : Fri Jan 03 16:07:20 2020
Response via : Initial Calibration

