

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW010920\
 Data File : VW014588.D
 Acq On : 09 Jan 2020 17:47
 Operator : SY/VA
 Sample : L1063-03
 Misc : 3.243G/5ML/MSVOA W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WP022SB451 200108 30-32

Quant Time: Jan 10 02:06:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jan 03 16:07:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	594299	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	922977	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	655143	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.55	152	222146	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	321646	61.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.80%	
35) Dibromofluoromethane	7.88	113	315019	59.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.44%	
50) Toluene-d8	10.32	98	1129616	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	12.62	95	302547	40.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.04%	
Target Compounds						
16) Acetone	4.12	43	116243	79.228	ug/l	100
18) Methyl Acetate	4.67	43	47055	13.279	ug/l	100
20) Methylene Chloride	4.91	84	66114	7.281	ug/l	97
27) cis-1,2-Dichloroethene	7.16	96	17287	2.620	ug/l	94
65) Chlorobenzene	11.65	112	74976	5.586	ug/l	98
88) 1,4-Dichlorobenzene	13.57	146	19843	2.730	ug/l #	57
91) 1,2-Dichlorobenzene	13.86	146	11549	1.785	ug/l #	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW010920\
Data File : VW014588.D
Acq On : 09 Jan 2020 17:47
Operator : SY/VA
Sample : L1063-03
Misc : 3.243G/5ML/MSVOA W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WP022SB451_200108_30-32

Quant Time: Jan 10 02:06:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Quant Title : SW846 8260
QLast Update : Fri Jan 03 16:07:20 2020
Response via : Initial Calibration

