

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011019\
 Data File : VW008184.D
 Acq On : 11 Jan 2019 08:40
 Operator : SY/AP
 Sample : K1102-03
 Misc : 5.07G/5ML/MSVOA W/SOIL
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CPSB-02(15-20)

Quant Time: Jan 11 11:27:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 04:40:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	89326	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	156410	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	142800	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	66850	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	50712	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
35) Dibromofluoromethane	7.88	113	48531	53.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.40%	
50) Toluene-d8	10.32	98	195604	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.62	95	67386	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
Target Compounds						
16) Acetone	4.14	43	9765	36.008	ug/l	98
17) Carbon Disulfide	4.37	76	3213	1.051	ug/l #	84
20) Methylene Chloride	4.92	84	7236	2.780	ug/l #	88
39) Methylcyclohexane	9.33	83	15496	7.585	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011019\
Data File : VW008184.D
Acq On : 11 Jan 2019 08:40
Operator : SY/AP
Sample : K1102-03
Misc : 5.07G/5ML/MSVOA W/SOIL
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CPSB-02(15-20)

Quant Time: Jan 11 11:27:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 04:40:48 2019
Response via : Initial Calibration

