


Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011020\
Data File : VW014594.D
Acq On : 10 Jan 2020 08:50
Operator : SY/VA
Sample : BFB
Misc : 5.00G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010320S.M
Title : SW846 8260
Last Update : Fri Jan 03 16:07:20 2020



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1747

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	60978	PASS
75	95	30	60	46.8	171029	PASS
95	95	100	100	100.0	365440	PASS
96	95	5	9	6.7	24490	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	82.4	300992	PASS
175	174	5	9	8.0	23933	PASS
176	174	95	101	98.6	296832	PASS
177	176	5	9	6.8	20186	PASS