

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW011119\
 Data File : VW008198.D
 Acq On : 11 Jan 2019 17:07
 Operator : SY/VA
 Sample : K1088-03
 Misc : 5.03G/5ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 CPSB-06(15-20)

Quant Time: Jan 12 00:37:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 09 04:40:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	90659	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	163672	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	153455	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	68081	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	56238	55.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.38%	
35) Dibromofluoromethane	7.88	113	52367	54.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.72%	
50) Toluene-d8	10.32	98	215581	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
62) 4-Bromofluorobenzene	12.62	95	74038	51.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.12%	
Target Compounds						
16) Acetone	4.15	43	16391	74.660	ug/l #	81
20) Methylene Chloride	4.91	84	8261	3.791	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW011119\
Data File : VW008198.D
Acq On : 11 Jan 2019 17:07
Operator : SY/VA
Sample : K1088-03
Misc : 5.03G/5ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
CPSB-06(15-20)

Quant Time: Jan 12 00:37:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W010819S.M
Quant Title : SW846 8260
QLast Update : Wed Jan 09 04:40:48 2019
Response via : Initial Calibration

