

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW011121\
 Data File : VW017815.D
 Acq On : 11 Jan 2021 13:41
 Operator : SY/VA
 Sample : M1074-03
 Misc : 6.16G/10ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFT42

Quant Time: Jan 12 05:56:04 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
 Quant Title : VOC Analysis
 QLast Update : Tue Jan 12 05:24:59 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	319560	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	298828	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	150115	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	105826	21.40	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.60%
7) Chloroethane-d5	2.89	69	81284	19.92	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.68%
10) 1,1-Dichloroethene-d2	4.02	63	136691	15.87	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.48%
20) 2-Butanone-d5	7.08	46	36483	45.53	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.06%
24) Chloroform-d	7.65	84	167921	20.11	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.44%
26) 1,2-Dichloroethane-d4	8.31	65	95413	22.08	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	88.32%
29) Benzene-d6	8.27	84	334896	19.89	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.56%
33) 1,2-Dichloropropane-d6	9.27	67	96037	20.70	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	82.80%
37) Toluene-d8	10.32	98	308874	19.11	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.44%
38) trans-1,3-Dichloropropene-	10.58	79	40187	19.43	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.72%
39) 2-Hexanone-d5	10.93	63	26242	45.04	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.08%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	70378	20.43	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.72%
61) 1,2-Dichlorobenzene-d4	13.85	152	96250	18.41	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	73.64%#

Target Compounds					Ovalue
13) Acetone	4.12	43	1951	2.819 ug/L	74
16) Methylene chloride	4.92	84	9200	2.013 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW011121\
Data File : VW017815.D
Acq On : 11 Jan 2021 13:41
Operator : SY/VA
Sample : M1074-03
Misc : 6.16G/10ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFT42

Quant Time: Jan 12 05:56:04 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
Quant Title : VOC Analysis
QLast Update : Tue Jan 12 05:24:59 2021
Response via : Initial Calibration

