

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011122\
 Data File : VW021847.D
 Acq On : 11 Jan 2022 21:20
 Operator : SY/VA
 Sample : VIBLK581
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK581

Quant Time: Jan 12 00:52:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jan 12 00:49:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	109208	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	111052	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	66677	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	55003	27.098	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 108.400%	
7) Chloroethane-d5	2.885	69	43074	23.500	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 94.000%	
11) 1,1-Dichloroethene-d2	4.013	63	62933	19.639	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	= 78.560%	
21) 2-Butanone-d5	7.092	46	17112	59.728	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 119.460%	
24) Chloroform-d	7.653	84	86789	24.600	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 98.400%	
26) 1,2-Dichloroethane-d4	8.305	65	45784	26.873	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 107.480%	
32) Benzene-d6	8.275	84	180137	24.999	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 100.000%	
36) 1,2-Dichloropropane-d6	9.274	67	46522	24.495	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 97.960%	
41) Toluene-d8	10.323	98	178906	24.629	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 98.520%	
43) trans-1,3-Dichloroprop...	10.579	79	18536	22.986	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 91.960%	
47) 2-Hexanone-d5	10.927	63	15943	61.863	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 123.720%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	40794	27.468	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 109.880%	
66) 1,2-Dichlorobenzene-d4	13.853	152	69887	25.491	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 101.960%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011122\
Data File : VW021847.D
Acq On : 11 Jan 2022 21:20
Operator : SY/VA
Sample : VIBLK581
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK581

Quant Time: Jan 12 00:52:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jan 12 00:49:08 2022
Response via : Initial Calibration

