

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011222\
 Data File : VW021858.D
 Acq On : 12 Jan 2022 11:31
 Operator : SY/VA
 Sample : N1084-07
 Misc : 6.18g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLN0

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/13/2022
 Supervised By :Mahesh Dadoda 01/13/2022

Quant Time: Jan 12 23:50:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jan 12 23:48:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	209990	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	170562	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	60156	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	50587	12.961	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	51.840%		
7) Chloroethane-d5	2.892	69	41543	11.787	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	47.160%		
11) 1,1-Dichloroethene-d2	4.025	63	72522	11.769	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	47.080%		
21) 2-Butanone-d5	7.098	46	23478	42.618	ug/L	0.01
Spiked Amount 50.000	Range 20 - 135		Recovery =	85.240%		
24) Chloroform-d	7.653	84	111287	16.404	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	65.600%		
26) 1,2-Dichloroethane-d4	8.311	65	56055	17.111	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	68.440%#		
32) Benzene-d6	8.275	84	224912	20.322	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	81.280%		
36) 1,2-Dichloropropane-d6	9.274	67	60123	20.611	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	82.440%		
41) Toluene-d8	10.323	98	211748	18.980	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	75.920%		
43) trans-1,3-Dichloroprop...	10.579	79	21625	17.460	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	69.840%		
47) 2-Hexanone-d5	10.927	63	18812	47.527	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	95.060%		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	39790	17.444	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	69.760%		
66) 1,2-Dichlorobenzene-d4	13.853	152	44355	17.932	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	71.720%#		
Target Compounds					Qvalue	
13) Acetone	4.172	43	1848m	4.060	ug/L	
19) 1,1-Dichloroethane	6.232	63	2782	0.560	ug/L	# 91
30) 1,1,1-Trichloroethane	7.878	97	4586	1.037	ug/L	# 92
46) Tetrachloroethene	10.860	164	36276	12.090	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011222\
Data File : VW021858.D
Acq On : 12 Jan 2022 11:31
Operator : SY/VA
Sample : N1084-07
Misc : 6.18g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLN0

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/13/2022
Supervised By : Mahesh Dadoda 01/13/2022

Quant Time: Jan 12 23:50:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jan 12 23:48:31 2022
Response via : Initial Calibration

