

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011222\
 Data File : VW021862.D
 Acq On : 12 Jan 2022 13:41
 Operator : SY/VA
 Sample : N1114-03
 Misc : 5.28g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BGLP5

Quant Time: Jan 12 23:50:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Jan 12 23:48:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	231912	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.634	117	228395	25.000	ug/L	# 0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	133346	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	56372	13.078	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	52.320%
7) Chloroethane-d5	2.885	69	42385	10.889	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	43.560%
11) 1,1-Dichloroethene-d2	4.019	63	85382	12.547	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	50.200%
21) 2-Butanone-d5	7.086	46	29625	48.693	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.380%
24) Chloroform-d	7.653	84	139342	18.598	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	74.400%
26) 1,2-Dichloroethane-d4	8.311	65	72565	20.057	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	80.240%
32) Benzene-d6	8.275	84	282374	19.054	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.200%
36) 1,2-Dichloropropane-d6	9.274	67	78236	20.029	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.120%
41) Toluene-d8	10.323	98	284135	19.019	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	76.080%
43) trans-1,3-Dichloroprop...	10.579	79	32630	19.675	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.680%
47) 2-Hexanone-d5	10.927	63	26702	50.378	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	100.760%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	63523	20.797	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.200%
66) 1,2-Dichlorobenzene-d4	13.853	152	115122	20.996	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	84.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW011222\
Data File : VW021862.D
Acq On : 12 Jan 2022 13:41
Operator : SY/VA
Sample : N1114-03
Misc : 5.28g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BGLP5

Quant Time: Jan 12 23:50:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM011122SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Jan 12 23:48:31 2022
Response via : Initial Calibration

