

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW011321\
 Data File : VW017864.D
 Acq On : 13 Jan 2021 21:47
 Operator : SY/VA
 Sample : M1099-05
 Misc : 4.70G/10ML/MSVOA W/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 F1C05

Quant Time: Jan 14 15:09:49 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 14 11:39:57 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	313533	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	247482	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.56	152	85838	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.35	65	128575	26.50	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.00%
7) Chloroethane-d5	2.89	69	106551	26.61	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.44%
10) 1,1-Dichloroethene-d2	4.02	63	160233	18.96	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.84%
20) 2-Butanone-d5	7.08	46	61539	78.27	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	156.54%#
24) Chloroform-d	7.65	84	215509	26.30	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	105.20%
26) 1,2-Dichloroethane-d4	8.31	65	134039	31.61	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	126.44%
29) Benzene-d6	8.27	84	417640	29.96	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	119.84%
33) 1,2-Dichloropropane-d6	9.27	67	127966	33.30	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	133.20%#
37) Toluene-d8	10.32	98	339295	25.34	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	101.36%
38) trans-1,3-Dichloropropene-	10.58	79	38221	22.31	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.24%
39) 2-Hexanone-d5	10.93	63	40082	83.06	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	166.12%#
48) 1,1,2,2-Tetrachloroethane-	12.69	84	95746	33.56	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	134.24%#
61) 1,2-Dichlorobenzene-d4	13.85	152	77601	25.96	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.84%

Target Compounds

					Ovalue
13) Acetone	4.13	43	8548	12.590	ug/L 96
14) Carbon disulfide	4.39	76	23422	2.030	ug/L 99
16) Methylene chloride	4.92	84	33792	7.537	ug/L 98
21) 2-Butanone	7.18	43	2200	2.409	ug/L 89
63) 1,3-Dichlorobenzene	13.50	146	9245	1.745	ug/L 93
64) 1,4-Dichlorobenzene	13.58	146	5898	1.108	ug/L 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW011321\
Data File : VW017864.D
Acq On : 13 Jan 2021 21:47
Operator : SY/VA
Sample : M1099-05
Misc : 4.70G/10ML/MSVOA W/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
F1C05

Quant Time: Jan 14 15:09:49 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 14 11:39:57 2021
Response via : Initial Calibration

