

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW011321\
 Data File : VW017863.D
 Acq On : 13 Jan 2021 21:24
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00G/10ML/MSVOA W/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK

Quant Time: Jan 14 15:01:50 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 14 11:39:57 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.84	114	323852	25.00	ug/L	0.00
28) Chlorobenzene-d5	11.63	117	302822	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.55	152	142001	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.36	65	110500	22.05	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.20%
7) Chloroethane-d5	2.89	69	90742	21.94	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.76%
10) 1,1-Dichloroethene-d2	4.02	63	145613	16.68	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	66.72%
20) 2-Butanone-d5	7.07	46	47626	58.64	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.28%
24) Chloroform-d	7.65	84	196487	23.22	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.88%
26) 1,2-Dichloroethane-d4	8.31	65	115799	26.44	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.76%
29) Benzene-d6	8.27	84	384286	22.53	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.12%
33) 1,2-Dichloropropane-d6	9.27	67	113748	24.19	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	96.76%
37) Toluene-d8	10.32	98	354343	21.63	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.52%
38) trans-1,3-Dichloropropene-	10.58	79	43167	20.60	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.40%
39) 2-Hexanone-d5	10.93	63	32143	54.44	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.88%
48) 1,1,2,2-Tetrachloroethane-	12.69	84	88638	25.39	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	101.56%
61) 1,2-Dichlorobenzene-d4	13.85	152	109080	22.06	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.24%


Target Compounds					Ovalue
16) Methylene chloride	4.92	84	7827	1.690 ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW011321\
Data File : VW017863.D
Acq On : 13 Jan 2021 21:24
Operator : SY/VA
Sample : VIBLK
Misc : 5.00G/10ML/MSVOA W/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK

Quant Time: Jan 14 15:01:50 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\SOM2WLM122420S.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 14 11:39:57 2021
Response via : Initial Calibration





#16
 Methvlene chloride
 Concen: 1.690 ug/L
 RT: 4.92 min Scan# 494
 Delta R.T. 0.00 min
 Lab File: VW017863.D
 Acq: 13 Jan 2021 21:24

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK

Tot Ion: 84 Resp: 7827
 Ion Ratio Lower Upper
 84 100
 49 127.9 90.5 168.1
 86 72.4 44.8 83.2

